

UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

DATE MAILED: 12/05/2006

APPLICATION NO.	F	ILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO	
10/665,480	09/22/2003		Reinhold Schmieding	A8130.0028/P028-A	5525	
24998	7590	12/05/2006		EXAMINER		
DICKSTEI 1825 EYE S				PRONE, CHRI	STOPHER D	
Washington, DC 20006-5403				ART UNIT	ART UNIT PAPER NUMBER	
_			·	3738		

Please find below and/or attached an Office communication concerning this application or proceeding.

•						
	Application No.	Applicant(s)				
Office Action Comments	10/665,480	SCHMIEDING, REINHOLD				
Office Action Summary	Examiner	Art Unit				
	Christopher D. Prone	3738				
The MAILING DATE of this communication app Period for Reply	ears on the cover sheet with the c	orrespondence address				
A SHORTENED STATUTORY PERIOD FOR REPLY WHICHEVER IS LONGER, FROM THE MAILING DA - Extensions of time may be available under the provisions of 37 CFR 1.13 after SIX (6) MONTHS from the mailing date of this communication. - If NO period for reply is specified above, the maximum statutory period w - Failure to reply within the set or extended period for reply will, by statute, Any reply received by the Office later than three months after the mailing earned patent term adjustment. See 37 CFR 1.704(b).	ATE OF THIS COMMUNICATION 36(a). In no event, however, may a reply be tim vill apply and will expire SIX (6) MONTHS from cause the application to become ABANDONE	N. nely filed the mailing date of this communication. D (35 U.S.C. § 133).				
Status						
1) Responsive to communication(s) filed on 28 Se	eptember 2006.					
,— ,—	This action is FINAL . 2b) ☐ This action is non-final.					
	Since this application is in condition for allowance except for formal matters, prosecution as to the merits is					
closed in accordance with the practice under E	x parte Quayle, 1935 C.D. 11, 45	i3 O.G. 213.				
Disposition of Claims		,				
4) Claim(s) 10-15 is/are pending in the application 4a) Of the above claim(s) is/are withdraw 5) Claim(s) is/are allowed. 6) Claim(s) 10-15 is/are rejected. 7) Claim(s) is/are objected to. 8) Claim(s) are subject to restriction and/or	vn from consideration.					
Application Papers						
9) The specification is objected to by the Examiner 10) The drawing(s) filed on is/are: a) access Applicant may not request that any objection to the of Replacement drawing sheet(s) including the correction in the original of the correction of the original of the correction of the original original original or the correction of the original origi	epted or b) objected to by the Edrawing(s) be held in abeyance. See on is required if the drawing(s) is obj	e 37 CFR 1.85(a). ´ ected to. See 37 CFR 1.121(d).				
Priority under 35 U.S.C. § 119	•	•				
 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of: 1. Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No. 3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). * See the attached detailed Office action for a list of the certified copies not received. 						
Attachment(s) 1) Notice of References Cited (PTO-892) 2) Notice of Draftsperson's Patent Drawing Review (PTO-948) 3) Information Disclosure Statement(s) (PTO/SB/08) Paper No(s)/Mail Date	4) Interview Summary Paper No(s)/Mail Da 5) Notice of Informal Pa 6) Other:	te				

Application/Control Number: 10/665,480

Art Unit: 3738

DETAILED ACTION

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(b) The invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claim 10, 11, and 13-15 are rejected under 35 U.S.C. 102(b) as being anticipated by International Application Publication WO 99/21515 Grooms.

In response to claims 10, 11, and 13-15 Grooms discloses the same invention being an implant shown in figure 2B comprising a blunt proximal end (202), a tapered distal end adapted to receive a suture (201), a transverse eye (210), and a cylindrical shaft (204). Grooms also teaches the use of channels formed on the opposite sides of the implant shown as the channels in figure 7D that may accommodate suture received in the eye of the implant.

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Application/Control Number: 10/665,480

Art Unit: 3738

Claim 12 is rejected under 35 U.S.C. 103 as being unpatentable over Grooms in view of United States Patent 6,045,554 Grooms et al.

Grooms discloses the invention substantially as claimed being an implant described above. However, Grooms does not disclose that the implant is made of allograft or a synthetic bone material.

Grooms 554' teaches the use of an implant made of allograft or synthetic bone material (2:9-22) in the same field of endeavor for the purpose of providing an implant that is capable of fusing with the bone it is implanted into.

It would have been obvious to one having ordinary skill in the art at the time the invention was made to combine the allograft or synthetic bone materials taught by Grooms 554' with the implant of Grooms in order to provide an implant that has strength comparable to a metallic version, and has the advantage of leaving no residual hardware while contributing to bone stock.

Response to Arguments

Applicant's argues that the shaft of Grooms is not a solid with a cylindrical shape. However, the shaft shown in figure 2B of Grooms is clearly cylindrical. The drawing is only wavy to show the flexibility of the shaft, but it still maintains a cylindrical cross-section. The examiner does not agree with the applicant's argument that since it is flexible it is not solid. The flexibility of an object has no relevance to it being a solid. From the disclosure of Brooms it is clear that the shaft 203 is a solid. The applicant further argues that the surface of Brooms

Application/Control Number: 10/665,480

Art Unit: 3738

implant is not smooth, but as is clear from the figures there is clearly a portion of the shaft 203 that has a smooth outer surface.

Conclusion

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Christopher D. Prone whose telephone number is (571) 272-6085. The examiner can normally be reached on Monday Through Fri 8:30 to 5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Corrine McDermott can be reached on (571) 272-4754. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3738

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

Christopher D Prone Examiner Art Unit 3738

₩ CDP

> CORRINE McDERMOTT SUPERVISORY PATENT EXAMINER TECHNOLOGY CENTER 3700